
PO2-1120 Percent Oxygen Sensor



Specifications:


Sensor Technology	Micro Fuel Cell
Signal Output ¹	28 - 52 uA
Measuring Range	0 - 100 Percent
Response Time T90	13 Seconds
Accuracy Full Scale ²	± 1% of Signal
Repeatability F.S.	± 0.5%
Temp Coefficient	2.54 % / ° C
Operating Temp	0 to 50 ° C
Storage Temp	0 to 45 ° C
Recommended Flow	0.5 - 2 SCFH
Humidity	0 - 100% RH (Non-Condensing)
Expected Life ³	120 Months
Storage	9 Months
Warranty ⁴	12 Months
PCB Connection	CenterFoilNegative Outer Foil Positive

- 1. Signal output is measured in air at 25 ° C, sea level.
- 2. Accuracy full scale is calculated with constant pressure and temperature & proper calibration (90% O2 value on full scale range). Drastic temperature change can result in a maximum error of ± 10%.
- 3. Expected life is calculated when O2 is 20.9% @ 25 ° C, sea level.
- 4. Southland Sensing warrants these sensors for the period noted above to be free from defects in materials and workmanship. Southland Sensing will not be held liable for sensors damaged due to customer neglect.





PCB Assembly



Direct Replacement Guide

Analytical Industries: GPR-11-120; GPR-11-120OP

深圳市新世联科技有限公司

地址:深圳市深南中路2066号华能大厦712室

电话: 0755-83680810 83680820 83680830 83680860

网址: www.apollounion.com

邮编: 518031 传真: 0755-83680866

邮箱: sales@apollounion.com